

MICHIGAN DEPT OF AGRICULTURE & RURAL DEVELOPMENT  
 LABORATORY DIVISION  
 WEIGHTS AND MEASURES PROGRAM  
 (517) 655 - 8202  
 michigan.gov/wminfo

Device Detail

**Insp Date:** 9/30/2015      **Business ID:** 91169  
**Business:** ENVIRO SOLIDS  
 6011 WYOMING  
 DEARBORN, MI 48126

**Inspection:** BB000845  
**Store ID:**  
**Phone:** (313)582-8032  
**Inspector:** 015 James Barnes  
**Reason:** NEW

**Make:** FAIRBANKS      **Type:** Heavy Capacity Scale      **Results:** Sealed  
**Model:** FB 2550      **Subtype:** 09-5 (25000)      **Other:**  
**Serial:** 140870050235      **Insp Type:** Maintenance      **Seal #:** AUDIT TRAIL  
**Loc:** RIGHT OF ENTRANCE  
**Elapsed Time:** 2.5      **Capacity:** 200,000      **Units:** lb  
**d:** 20      **Class:** III L      **Sections:** 5  
**CLC or Section Capacity:** 90,000      **NTEP:** Yes      **Weighing Elements:** Electronic Load  
 Cel  
**Indicating Elements:** DWI      **Recording Elements:** CPU Interface      **Deck Size:** 11' X 80'  
**Invoice#:** N/A

Type	Name	Inc/Base	Actual	Display	Err	Tol	Res	Notes
Test	Zero Load	0	0	0	0	20	Pass	
Test	Increasing Load	5000	5000	5000	0	20	Pass	Section 1
Test	Increasing Load	4000	9000	9000	0	20	Pass	
Test	Increasing Load	4000	13000	13000	0	40	Pass	
Test	Shift	0	13000	13000	0	40	Pass	Section 2
Test	Shift	0	13000	13000	0	40	Pass	Section 3
Test	Shift	0	13000	13000	0	40	Pass	Section 4
Test	Shift	0	13000	13000	0	40	Pass	Section 5
Test	Range(Shift)	0	13000		0	40	Pass	
Test	Increasing Load	4000	17000	17000	0	40	Pass	Section 1
Test	Increasing Load	4000	21000	21000	0	60	Pass	
Test	Increasing Load	4000	25000	25000	0	60	Pass	
Test	Shift	0	25000	25000	0	60	Pass	Midspan
Test	Shift	0	25000	25000	0	60	Pass	Section 2
Test	Shift	0	25000	25000	0	60	Pass	Midspan
Test	Shift	0	25000	25000	0	60	Pass	Section 3
Test	Shift	0	25000	25000	0	60	Pass	Midspan
Test	Shift	0	25000	25000	0	60	Pass	Section 4
Test	Shift	0	25000	25000	0	60	Pass	Midspan
Test	Shift	0	25000	25020	20	60	Pass	Section 5
Test	Range(Shift)	0	25000		20	60	Pass	
Test	Decreasing Load	-12000	13000	13000	0	40	Pass	Section 1
Test	Zero Load	0	0	0	0	20	Pass	
Test	Strain	27940	25000	52960	20	60	Pass	

  
 Inspector

Acknowledged Receipt: \_\_\_\_\_

### Device Detail

**Notes:**

Reviewed place in service reports and test reports from Grand Rapids Metrology dated 11-12-14 and 1-12-15. Scale was newly installed on 11-12-14 by Grand Rapids Metrology and placed into service. Verified place in service reports and test reports were sent into the MDARD Weights & Measures office.

This is the first inspection by MDARD performed on this scale.

**AUDIT TRAIL**

Calibration Date: Never  
Calibration Count: 1  
Configuration Date: 1-12-15  
Configuration Count: 2

**INDICATOR**

Make: Fairbanks  
Model: FB 2550  
S/N: 140870050235  
COC: 10-089  
Class: III / IIL  
nMAX: 10,000

**WEIGHBRIDGE**

Make: Thurman  
Model: 8560  
S/N: 61998  
COC: 89-016  
Class: IIL  
nMAX: 10,000  
eMIN: 20 lbs  
CLC: 90,000  
Capacity: 200,000 lbs

  
Inspector

\_\_\_\_\_  
Acknowledged Receipt: